

Special 510(k) Summary

MAY 24 2012

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 872.1800.

Date

4/18/2012

Manufacturer

Vatech Co., Ltd.

23-4, Seogu-Dong, Hwaseong-Si, Gyeonggi-Do, 445-170, Korea Republic

Tel: +82-31-379-9585

Fax: +82-31-379-9638

Contact person: Mr. Sonny Park

Official Correspondent (U.S. Designated agent)

Mtech Group

12946 Kimberley Ln, Houston, TX 77079

Tel: +713-467-2607

Fax: +713-464-8880

Contact person: Mr. Dave Kim (davekim@mtech-inc.net)

Trade/Proprietary Name:

PaX-Duo3D Plus (PCT-5000)

Common Name:

Computed Tomography X-ray System

Classification Name:

X-ray, Tomography, Computed, Dental (21CFR 892.1750, Product code OAS, Class2)

Description:

PaX-Duo3D Plus (PCT-5000), a dental radiographic imaging system, consists of dual image acquisition modes; panorama, and cone beam computed tomography. Specifically designed for dental radiography of the teeth or jaws, PaX-Duo3D Plus (PCT-5000) is a complete dental X-ray system equipped with x-ray tube, generator and dedicated SSXI detector for dental panoramic and cone beam computed tomographic radiography.

The dental CBCT system is based on CMOS digital X-ray detector. CMOS CT detector is used to capture radiographic diagnostic images of oral anatomy in 3D for dental treatment such as oral surgery or implant. The device can also be operated as the panoramic dental x-ray system based on CMOS X-ray detector.

Indication for use:

PaX-Duo3D Plus is a computed tomography x-ray system intended to take panoramic and cross-sectional images of the oral and craniofacial anatomy to provide diagnostic information for adult and pediatric patients. The device is operated and used by physicians, dentists, dental assistants, x-ray technicians and other professionals who are licensed by the law of the State in which he or she practices to use the device.

Predicate Device:

Manufacturer	: Vatech Co., Ltd
Device	: PaX-Duo3D Plus
510(k) Number	: K102102 (Decision Date – 3/11/2011)

Substantial Equivalence:

PaX-Duo3D Plus (PCT-5000) described in this 510(k) has the similar intended use and technical characteristics as PaX-Duo3D Plus of Vatech Co., Ltd.

Characteristic	Proposed Vatech Co., Ltd. PaX-Duo3D Plus (PCT-5000)	Predicate Vatech Co., Ltd. PaX-Duo3D Plus
-----------------------	--	--

510(k) number	-	K102102
Indications for use	PaX-Duo3D Plus is a computed tomography x-ray system intended to take panoramic and cross-sectional images of the oral and craniofacial anatomy to provide diagnostic information for adult and pediatric patients. The device is operated and used by physicians, dentists, dental assistants, x-ray technicians and other professionals who are licensed by the law of the State in which he or she practices to use the device.	PaX-Duo3D Plus is a computed tomography x-ray system intended to take panoramic and cross-sectional images of the oral and craniofacial anatomy to provide diagnostic information for adult and pediatric patients. The device is operated and used by physicians, dentists, dental assistants, x-ray technicians and other professionals who are licensed by the law of the State in which he or she practices to use the device.
Performance Specification	Panoramic and computed tomography (CBCT)	Panoramic and computed tomography (CBCT)
Input Voltage	AC 100-120 / 200-240 V	AC 110 / 230 V ~
Tube Voltage	50-90 kV	50-90 kV
Tube Current	2-10 mA	2-10 mA
Focal Spot Size	0.5 mm	0.5 mm
Exposure Time	Max 20.2 s	6.8 - 24 s
Total Filtration	2.8 mmAl	2.8 mmAl
Size of Imaging Volume	CT (Xmaru0712CF) : 5 x 5 cm / 8 x 5 cm / 8 x 8 cm CT (Xmaru1215CF Plus) : 5 x 5 cm / 8 x 5 cm / 8.5 x 8.5 cm / 12 x 8.5 cm CT(Xmaru1215CF Master Plus) : 5 x 5 cm / 8 x 5 cm / 8.5 x 8.5 cm / 12 x 8.5 cm CT(Xmaru1524CF) : 5 x 5 cm / 8 x 5 cm / 8.5 x 8.5 cm / 12 x 8.5 cm	CT(Xmaru1524CF) : 5 x 5 cm / 8 x 5 cm / 8.5 x 8.5 cm / 12 x 8.5 cm

Pixel Resolution	Panoramic (Xmaru1501CF) : 5.0 lp/mm	Panoramic (Xmaru1501CF) : 5.0 lp/mm
	CT (Xmaru0712CF) : 3.5 lp/mm	CT (Xmaru1524CF) : 2.5 lp/mm
	CT (Xmaru1215CF Plus) : 3.5 lp/mm	
	CT(Xmaru1215CF Master Plus) : 10.1 lp/mm- full resolution 5.0 lp/mm -2x2 binning 2.5 lp/mm-4x4 binning	
	CT(Xmaru1524CF) : 2.5 lp/mm	
Pixel Size	Panoramic (Xmaru1501CF) : 100 μ m	Panoramic (Xmaru1501CF) : 100 μ m
	CT (Xmaru0712CF) : 140 μ m	CT (Xmaru1524CF) : 200 μ m
	CT (Xmaru1215CF Plus) : 140 μ m	
	CT(Xmaru1215CF Master Plus) : 49.5 μ m-full resolution 99 μ m-2x2 binning 198 μ m-4x4 binning	
	CT(Xmaru1524CF) : 200 μ m	
Image Receptor	CMOS photodiode array – panoramic (Xmaru1501CF) & CT(Xmaru0712CF, Xmaru1215CF Plus, Xmaru1215CF Master Plus, Xmaru1524CF)	CT (Xmaru1524CF) with Flat Panel Detector

Indications for use, safety characteristics, and non-clinical performance for panoramic and CBCT sensors of PaX-Duo3D Plus (PCT-5000) and PaX-Duo3D Plus are similar. The primary differences are as follows: PaX-Duo3D Plus (PCT-5000) introduces three new cone beam CT sensors: Xmaru0712CF, Xmaru1215CF Plus, Xmaru1215CF Master Plus. The non-clinical performance and clinical consideration report for the new SSXI CBCT sensors are provided separately in this submission. Based on the non-clinical and clinical consideration and the outcome of an expert review of image comparisons for both devices, new PaX-Duo3D Plus (PCT-5000) is substantially equivalent, in terms of safety and effectiveness, with PaX-Duo3D Plus, the predicate device.

Safety, EMC and Performance Data:

Electrical, mechanical, environmental safety and performance testing according to standard IEC 60601-1(A1+A2, 1995), IEC 60601-1-1 (2001), IEC 60601-1-3 (Ed. 1, 1994), IEC 60601-2-7 (1998), IEC 60601-2-28 (Ed. 1, 1993), IEC 60601-2-32 (Ed. 1, 1994) and IEC 60601-2-44 (Ed. 2, 2002) were performed, and EMC testing were conducted in accordance with standard IEC 60601-1-2.

PaX-Duo3D Plus (PCT-5000) meets the provisions of NEMA PS 3.1-3.18, Digital Imaging and Communications in Medicine (DICOM) Set.

Non-clinical & Clinical considerations according to FDA Guidance “Guidance for the submissions of 510(k)’s for Solid State X-ray Imaging Devices” were performed.

Acceptance test according to IEC 61223-3-4 and IEC 61223-3-5 was performed.

All test results were satisfactory.

Conclusion:

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on

the information provided in this premarket notification. Vatech Co., Ltd. concludes that PaX-Duo3D Plus (PCT-5000) is safe and effective and substantially equivalent to predicate device as described herein.

END



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-G609
Silver Spring, MD 20993-0002

Vatech Co., Ltd.
% Mr. Dave Kim
Medical Device Regulatory Affairs
Mtech Group
12946 Kimberly Lane
HOUSTO TX 77079

MAY 24 2012

Re: K121236

Trade/Device Name: PaX-Duo3D Plus
Regulation Number: 21 CFR 892.1750
Regulation Name: Computed tomography x-ray system
Regulatory Class: II
Product Code: OAS
Dated: April 18, 2012
Received: April 24, 2012

Dear Mr. Kim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely Yours,

A handwritten signature in black ink, appearing to read "Janine M. Morris", is written over the typed name and title.

Janine M. Morris
Acting Director
Division of Radiological Devices
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(K) Number (if known): K121236

Device Name: PaX-Duo3D Plus

Classification: Computed tomography X-ray system

Indications for Use:


PaX-Duo3D Plus is a computed tomography x-ray system intended to take panoramic and cross-sectional images of the oral and craniofacial anatomy to provide diagnostic information for adult and pediatric patients. The device is operated and used by physicians, dentists, dental assistants, x-ray technicians and other professionals who are licensed by the law of the State in which he or she practices to use the device.

Prescription Use ☒
 (Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
 (Part 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety
 510K K121236

Concurrence of CDRH, Office of Device Evaluation(ODE)